



# VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130571

(CONTINUED FROM FRONT PAGE)

**C** TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

## DIAGNOSTIC CHECK

**Breath Test Record**

Serial Number 130571 Model DataMaster  
 Location ANC1 Anchorage Police Dept.

**Subject Information**

Last Name Martin O.L. # 1234567  
 First Name Patrick E

**Operator Information**

Operator Name Martin, Patrick E Case VOC  
 Operator # 4072 Department ANC1

**Test Results**

Alco Target .080 Alco Serial ZV  
 Nominal 0.078 Barometer 29.20 Test Type V  
 Dry Gas 1 0.079 Time: 07:21:25 Test Date 01/04/2011  
 Subject Alcohol 0.000 Time: 07:22:59 Test Time 07:20:13  
 Dry Gas 2 0.080 Time: 07:23:54 Report Y

**Diagnostic Test Record**

Serial Number 130571 Model DataMaster  
 Location ANC1 Anchorage Police Dept.

**Diagnostic Test Results**

Test Date 01/04/2011 Computer Okay  
 Test Time 07:16:31 Program Okay  
 Software Date 12/11/2008  
 Sample Chamber (C) 51  
 Breath Tube (C) 40  
 Barometer (in) 29.20  
 Flow Detector Okay  
 High Speed Pump Okay  
 Detector Okay  
 Filter Okay  
 Quartz Okay  
 Calibration Okay

*AMB 3/31/11*

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 29.20 in		
ALCO TARGET	.078	07:20
BLANK TEST	.000	07:21
INTERNAL STANDARD	.079	07:21
ALCO TV 29.20 in	.080	07:22
BLANK TEST	.000	07:22
SUBJECT SAMPLE	.000	07:22
BLANK TEST	.000	07:23
ALCO TV 29.20 in	.080	07:23
BLANK TEST	.000	07:24

HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY
PRINTER TEST	
1#432010144.../0120456789:1(=)P0AFCDEF6	